



# BioSim™ anti-Ipilimumab (Yervoy®) (Human) ELISA Kit

rev 07/18

(Catalog # E4398-100, 100 assays, Store at 4°C)

## I. Introduction:

Ipilimumab (Yervoy®) is a fully human IgG1κ antibody that binds to CTLA-4 (cytotoxic T lymphocyte associated antigen 4), a molecule on T-cells that is indicated for unresectable or metastatic melanoma. The absence or presence of CTLA-4 can augment or suppress the immune system's T-cell response in fighting disease. Ipilimumab is designed to block the activity of CTLA-4, thereby sustaining an active immune response in its attack on cancer cells. The proposed mechanism of action is indirect, and may be through T-cell - mediated anti-tumor immune responses. However, some patients develop unwanted immunogenicity, which leads to production of anti-drug-antibodies (ADAs) inactivating the therapeutic effects of the treatment and, in rare cases, inducing adverse effects. BioVision's BioSim™ anti-Ipilimumab ELISA kit is designed to detect antibody against Ipilimumab with high specificity and sensitivity in biological matrices.

## II. Application:

This ELISA kit is used for *in vitro* qualitative determination of antibody against Ipilimumab in serum and plasma

Cross Reactivity: Ipilimumab (Yervoy®) infusion camouflages/masks the presence of antibody to Ipilimumab (ATP) in serum/plasma samples. Therefore, blood sampling time is critical for detection of ATP. It is convenient to obtain blood sample just before the infusion or at least 2 weeks after the infusion of Ipilimumab.

## III. Sample Type:

Human serum and plasma

## IV. Kit Contents:

Components	E4398-100	Part No.
Micro ELISA Plate	1 plate	E4398-100-1
Positive Control	0.25 ml	E4398-100-2
Negative Control	0.5 ml	E4398-100-3
Assay Buffer	12 ml	E4398-100-4
Peroxidase Conjugate	12 ml	E4398-100-5
TMB substrate (Avoid light)	12 ml	E4398-100-6
Stop Solution	12 ml	E4398-100-7
Wash buffer (20X)	50 ml	E4398-100-8
Plate sealers	2	E4398-100-9

## V. User Supplied Reagents and Equipment:

- Microplate reader capable of measuring absorbance at 450 nm
- Precision pipettes with disposable tips
- Clean eppendorf tubes for preparing standards or sample dilutions
- Absorbent paper

## VI. Storage and Handling:

The entire kit may be stored at 4°C for up to 12 months from the date of shipment.

## VII. Reagent and Sample Preparation:

Note: Before using the kit, spin tubes and bring down all components to the bottom of tubes.

1. **Wash Buffer:** Dilute the 20X Wash Buffer to 1X solution in ddH<sub>2</sub>O (10 ml of Wash Buffer stock to 190 ml of ddH<sub>2</sub>O). Mix the 1X solution thoroughly by vortex manually. The working stock can be stable for 2 weeks after preparation at 4°C.
2. **Samples preparation:** The usual precautions for venipuncture should be observed. Samples are stable at 4°C for 7 days and -20°C for 6 months. Avoid freeze-and-thaw cycle.

## VIII. Assay Protocol:

Note: Bring all reagents, microplate and samples to room temperature 15 minutes prior to the assay.

It is recommended that all standards and samples be run at least in duplicate.

A standard curve must be run with each assay.

1. Prepare all reagents, samples and standards as instructed in section VII.
2. Pipette 100 µl of **Assay Buffer** into each of the wells to be used.
3. Add 10 µl of **negative control** (2 wells), **positive control**, and **samples** into appropriate wells. Cover wells and incubate for 60 minutes at room temperature (RT).
4. Discard incubation solution. Wash plate 3 times each with 300 µl of diluted **Wash Buffer**. Remove excess solution by tapping the inverted plate on a paper towel.
5. Add 100 µl of **Peroxidase Conjugate** into each well. Cover wells with adhesive plate sealer and incubate at RT for 60 minutes.
6. Discard the solution and wash the wells as step 3.
7. Add 100 µl of 1X **TMB substrate** solution and incubate the plate in dark at RT for 20 minutes
8. Add 100 µl of **Stop solution** to stop the reaction

**FOR RESEARCH USE ONLY! Not to be used on humans.**



9. Read the absorbance in micro plate reader set to 450 nm within 20 minutes. (reference wavelength to 650 nm)

#### IX. QUALITATIVE INTERPRETATION:

- For the run to be valid, the OD<sub>450/650</sub> nm of Positive Control (Standard A) should be  $\geq 1.0$  and the OD<sub>450/650</sub> nm of each Negative Control should be  $< 0.2$ , if not, improper technique or reagent deterioration may be suspected and the run should be repeated.
- If "Sample OD<sub>450/650</sub> / Negative Control OD<sub>450/650</sub>" is  $< 3$ , the sample is NEGATIVE for Antibody to Ipilimumab.
- If "Sample OD<sub>450/650</sub> / Negative Control OD<sub>450/650</sub>" is  $\geq 3$ , the sample is POSITIVE for Antibody to Ipilimumab.

#### X. RELATED PRODUCTS:

- BioSim™ Rituximab (Mabthera®) (Human) ELISA Kit (Cat. No. E4371-100)
- BioSim™ Adalimumab (Humira®) (Human) ELISA Kit (Cat. No. E4372-100)
- BioSim™ Bevacizumab (Avastin®) (Human) ELISA Kit (Cat. No. E4373-100)
- BioSim™ Etanercept (Enbrel®) (Human) ELISA Kit (Cat. No. E4374-100)
- BioSim™ Infliximab (Remicade®) (Human) ELISA Kit (Cat. No. E4375-100)
- BioSim™ Trastuzumab (Herceptin®)(Human) ELISA Kit (Cat. No. E4376-100)
- BioSim™ Golimumab (Simponi®)(Human) ELISA Kit (Cat. No. E4377-100)
- BioSim™ Infliximab (Remsima®)(Human) ELISA Kit (Cat. No. E4378-100)
- BioSim™ Cetuximab (Erbitux®)(Human) ELISA Kit (Cat. No. E4379-100)
- BioSim™ Denosumab (Prolia®)(Human) ELISA Kit (Cat. No. E4380-100)
- BioSim™ Omalizumab (Xolair®)(Human) ELISA Kit (Cat. No. E4381-100)
- BioSim™ Nivolumab (Opdivo®)(Human) ELISA Kit (Cat. No. E4382-100)
- BioSim™ Pembrolizumab (Keytruda®)(Human) ELISA Kit (Cat. No. E4383-100)
- BioSim™ Ipilimumab (Yervoy®)(Human) ELISA Kit (Cat. No. E4384-100)
- BioSim™ Rituximab (Mabthera®) (Human) ELISA Kit (Cat. No. E4385-100)
- BioSim™ Trastuzumab (Herceptin®) (Human) ELISA Kit (Cat. No. E4386-100)
- BioSim™ Infliximab (Remicade®) (Human) ELISA Kit (Cat. No. E4387-100)
- BioSim™ Adalimumab (Humira®) (Human) ELISA Kit (Cat. No. E4388-100)
- BioSim™ Bevacizumab (Avastin®) (Human) ELISA Kit (Cat. No. E4389-100)
- BioSim™ Infliximab (Remsima®) (Human) ELISA Kit (Cat. No. E4390-100)
- BioSim™ Cetuximab (Erbitux®) (Human) ELISA Kit (Cat. No. E4391-100)
- BioSim™ Etanercept (Enbrel®) (Human) ELISA Kit (Cat. No. E4392-100)
- BioSim™ Golimumab (Simponi®) (Human) ELISA Kit (Cat. No. E4393-100)
- BioSim™ Denosumab (Prolia®) (Human) ELISA Kit (Cat. No. E4394-100)
- BioSim™ Omalizumab (Xolair®) (Human) ELISA Kit (Cat. No. E4395-100)
- BioSim™ Nivolumab (Opdivo®) (Human) ELISA Kit (Cat. No. E4396-100)
- BioSim™ Pembrolizumab (Keytruda®) (Human) ELISA Kit (Cat. No. E4397-100)
- BioSim™ Ipilimumab (Yervoy®) (Human) ELISA Kit (Cat. No. E4398-100)
- BioSim™ Filgrastim (Herceptin®) (Human) ELISA Kit (Cat. No. E4399-100)