



09/19

# Prostacyclin ELISA Kit

(Catalog # E4767-100; 96 assays, Storage at 4°C)

## I. Introduction:

Prostacyclin is produced by endothelial and vascular muscle cells. It is available as a synthetic analogue, epoprostenol. Prostacyclin has potent and fast-acting antiplatelet effects but causes vasodilation at a higher dose. Prostacyclin has a short half-life, with vasodilator and platelet effects of about 30 minutes. Prostacyclin (PGI2) is involved in platelet aggregation, vasoconstriction, and reproductive functions. BioVision's Prostacyclin (PGI2) ELISA kit is based on the principle of Competitive-ELISA. The micro ELISA plate provided in this kit has been pre-coated with PGI2. During the reaction, PGI2 in the sample or standard competes with a fixed amount of PGI2 on the solid phase supporter for sites on the Biotinylated Detection Antibody specific to PGI2. Excess conjugate and unbound sample or standard are washed from the plate, and Avidin conjugated to Horseradish Peroxidase (HRP) are added to each microplate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color change is measured spectrophotometrically at a wavelength of 450 nm ± 2 nm. The concentration of PGI2 in the samples is then determined by comparing the OD of the samples to the standard curve.

# II. Applications:

In vitro, quantitative determination of human Osteocalcin

Detection Range: 31.25-2000 pg/ml

Sensitivity: 18.75 pg/ml

Precision: Coefficient of variation is < 10%.

## III. Sample Type:

Serum, plasma, other biological fluids

## IV. Kit Contents:

Components	E4767-100	Part Number	Storage Temp	
Micro ELISA Plate	8 wells x12 strips	E4767-100-1	-20°C	
Reference Standard	2 vials	E4767-100-2	-20°C	
Biotinylated Detection Antibody (100x)	120 µL	E4767-100-3	-20°C	
HRP Conjugate (100x)	120 µL	E4767-100-4	-20°C (Away from light)	
Reference Standard and Sample Diluent	20 ml	E4767-100-5	4°C	
Biotinylated Detection Antibody Diluent	14 ml	E4767-100-6	4°C	
HRP Conjugate Diluent	14 ml	E4767-100-7	4°C	
Wash Buffer (25x)	30 ml	E4767-100-8	4°C	
Substrate	10 ml	E4767-100-9	4°C (Away from light)	
Stop Solution	10 ml	E4767-100-10	4°C	
Plate Sealer	4	E4767-100-11	RT	

## V. User Supplied Reagents and Equipment:

- Microplate reader capable of measuring absorbance at 450 nm and 650 nm
- Precision pipettes with disposable tips
- · Clean eppendorf tubes for preparing standards or sample dilutions


## VI. Storage and Handling:

An unopened kit can be stored at 4  $^{
m o}$ C for 1 month. If the kit is not used within 1 month, store the items separately according to the following conditions once the kit is received.

## VII. Reagent and Sample Preparation:

Bring all reagents to room temperature before use. Before using the kit, spin tubes and bring down all components to the bottom of tubes.

- Wash Buffer (25x): Dilute 30 mL of Concentrated Wash Buffer with 720 mL of deionized or distilled water to prepare 750 mL of Wash Buffer. Note: if crystals have formed in the concentrate, warm it in a 40°C water bath and mix it gently until the crystals have completely dissolved.
- Biotinylated Detection Antibody: Calculate the required amount before the experiment (100 μl/well). Centrifuge the stock tube before use, dilute the 100x Concentrated Biotinylated Detection Ab to 1x working solution with Biotinylated Detection Antibody Diluent.
- HRP Conjugate: Calculate the required amount before the experiment (100µl/well). Dilute the 100x Concentrated HRP Conjugate to 1x working solution with HRP Conjugate Diluent.
- Standards: Centrifuge the standard at 10,000xg for 1 min. Add 1.0 mL of Standard and Sample Diluent, let it stand for 10 min and invert it gently several times. After it dissolves fully, mix it thoroughly with a pipette. This reconstitution produces a working solution of 2000 pg/ml. Then make serial dilutions as needed. The recommended dilution gradient is: 2000,





## **Gentaur Europe BVBA** Voortstraat 49, 1910 Kampenhout BELGIUM Tel 0032 16 58 90 45 info@gentaur.com



1000, 500, 250, 125, 62.5, 31.25, 0 pg/ml. Prepare 7 tubes, add 500  $\mu$ l of Standard & Sample Diluent to each tube. Pipette 500  $\mu$ l of the 2000 pg/ml stock solution to the first tube and mix up to produce a 1000 pg/ml working solution. Transfer 500  $\mu$ l of the solution into the other tube to form 2-fold serial dilutions of the highest standards to make the standard curve within the range of this assay.

# VIII. Sample Preparation:

**Note:** Samples should be assayed within 7 days when stored at 4°C, otherwise aliquot and stored at -20°C (≤1 month) or -80°C (≤3 months). Avoid repeated freeze-thaw cycles.

#### Serum

Allow samples to clot for 2 hours at room temperature or overnight at 4°C before centrifugation for 20 min at 1000×g at 2~8°C. Collect the supernatant to carry out the assay. Blood collection tubes should be disposable and endotoxin free.

#### Plasma

Collect plasma using EDTA-Na<sub>2</sub> as anticoagulant. Centrifuge samples for 15 min at 1000x g at 4°C within 30 min of collection. Collect the supernatant to carry out the assay. Hemolysed samples are not suitable for ELISA assay.

## · Cell lysates:

For adherent cells, gently wash the cells with moderate amount of pre-cooled PBS and dissociate the cells using trypsin. Collect the cell suspension into a centrifuge tube and centrifuge for 5 min at 1000×g. Discard the medium and wash the cells 3 times with pre-cooled PBS. For each 1×10<sup>6</sup> cells, add 150-250 µl of pre-cooled PBS to keep the cells suspended. Repeat the freeze-thaw process several times until the cells are fully lysed. Centrifuge for 10 min at 1500xg at 4°C. Remove the cell fragments, collect the supernatant for assay. Avoid repeated freeze-thaw cycles.

# <u>Tissue homogenates:</u>

It is recommended to get detailed references from the literature before analyzing different tissue types. For general information, hemolysed blood may affect the results, so the tissues should be minced into small pieces and rinsed in ice-cold PBS (0.01M, pH=7.4) to remove excess blood thoroughly. Tissue pieces should be weighed and then homogenized in PBS (tissue weight (g): PBS (mL) volume=1:9) with a glass homogenizer on ice. To further break down the cells, you can sonicate the suspension with an ultrasonic cell disrupter or subject it to freeze-thaw cycles. The homogenates are then centrifuged for 5 min at 5000xg to get the supernatant.

• <u>Cell culture supernatant or other biological fluids:</u> Centrifuge samples for 20 min at 1000xg at 2~ 8°C. Collect the supernatant for assay

## IX. Assay Protocol:

<u>Note:</u> Bring all reagents and samples to room temperature 30 minutes prior to the assay. It is recommended that all standards and samples be run at least in duplicate. A standard curve must be run with each assay.


- 1. Add 50 µl of each standard or samples into appropriate wells.
- 2. Immediately add 50 µl of **Biotinylated Detection Antibody** working solution to each well. Cover the plate with the sealer provided in the kit. Gently mix and incubate for 45 min. at 37°C. **Note:** solutions should be added to the bottom of the micro ELISA plate well, avoid touching the inside wall and bubble formation as much as possible.
- 3. Aspirate the solution from each well add 350 µl of 1x wash buffer to each well. Soak for 1~2 min and aspirate or decant the solution from each well and pat it dry against clean absorbent paper. Repeat this wash step 3 times.
  - Note: a microplate washer can be used in this step and other wash steps.
- 4. Add 100 µl of HRP Conjugate working solution to each well. Cover with the Plate sealer. Incubate for 30 min at 37°C.
- 5. Aspirate the solution from each well, repeat the wash process for five times as conducted in step 3.
- 6. Add 90 μl of **Substrate** to each well. Cover with a new plate sealer. Incubate for about 15 min at 37°C. Protect the plate from light. Note: the reaction time can be shortened or extended according to the actual color change, but not more than 30 min.
- Add 50 µl of Stop Solution to each well. Note: adding the stop solution should be done in the same order as the substrate solution.
- 8. Read the absorbance in micro plate reader set to 450 nm.

# X. Calculation:

Determine the average of the duplicate readings for each standard and samples then subtract the average zero standard optical density. Plot a four-parameter logistic with standard concentration on the x-axis and OD values on the y-axis. If the samples have been diluted, the

concentration calculated from the standard curve must be multiplied by the dilution factor. If the OD of the sample is under the lowest limit of the standard curve, retest the samples with appropriate dilution. The actual concentration is the concentration obtained by calculated multiplied by the dilution factor.

Typical standard curve and data is provided below for reference only. A standard curve must be run with each assay





## **Gentaur Europe BVBA** Voortstraat 49, 1910 Kampenhout BELGIUM Tel 0032 16 58 90 45 <u>info@gentaur.com</u>



Concentration(pg/mL)	2000	1000	500	250	125	62.5	31.25	0
OD	0.39	0.488	0.663	0.934	1.279	1.621	1.886	2.249

## Precision:

Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Human S100B were tested 20 times on one plate, respectively.

Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Human S100B were tested on 3 different plates, 20 replicates in each plate.

	Intra-assay	Intra-assay Precision			Inter-assay Precision		
Sample	1	2	3	1	2	3	
n	20	20	20	20	20	20	
Mean (pg/mL)	94.64	246.44	963.63	90.56	267.93	1038.28	
Standard deviation	5.82	11.68	44.91	5.80	15.86	39.56	
C V (%)	6.15	4.74	4.66	6.40	5.92	3.81	

## Recovery

The recovery of Human S100B spiked at three different levels in samples throughout the range of the assay was evaluated in various matrices.

Sample Type	Range (%)	Average Recovery (%)
Serum (n=5)	90-102	95
EDTA plasma (n=5)	91-102	96
Cell culture media (n=5)	85-96	91

# Linearity

Samples were spiked with high concentrations of Human S100B and diluted with Reference Standard & Sample Diluent to produce samples with values within the range of the assay.

		Serum (n=5)	EDTA plasma(n=5)	Cell culture media(n=5)
1:2	Range (%)	97-115	88-101	93-109
	Average (%)	105	94	101
1:4	Range (%)	94-108	86-98	94-107
	Average (%)	101	91	100
1:8	Range (%)	101-114	90-102	92-105
	Average (%)	108	96	99
1:16	Range (%)	100-115	90-103	92-110
	Average (%)	106	96	100

# **XI. RELATED PRODUCTS:**

- Prostaglandin D2 (PGD2) ELISA Kit (E4718)
- Prostaglandin E2 (PGE2) ELISA Kit Kit (E4637)
- Prostaglandin E1 (PGE1) ELISA Kit (E4716)

FOR RESEARCH USE ONLY! Not to be used on humans.