



BioSim™ anti-Evolocumab (Human) ELISA Kit

(Catalog # E5027-100, 96 assays, Store at 4°C)

I. Introduction:

Evolocumab is a human IgG2 monoclonal antibody that targets proprotein convertase subtilisin/kexin type 9 (PCSK9). PCSK9 is a serine protease that binds to LDL receptors on the hepatocyte cell surface and promotes the lysosome-mediated degradation of these receptors. This results in decreased uptake and clearance of LDL-C (bad cholesterol) by the receptors. The monoclonal antibody binds to PCSK9 and prevents interaction with the LDL receptor. This promotes the recycling of the internalized LDL receptor to the hepatocyte cell surface and increased removal of LDL-C from the blood. The original drug received approval from the FDA to treat homozygous and heterozygous familial cholesterolemia. The drug has also been approved to treat patients with high cholesterol, whose LDL-C was not controlled by the statin treatment. However, upon administration of the antibody, patients tend to develop an immune response that results in the formation of anti-Evolocumab antibodies, thus adversely affecting the therapeutic response to treatment. BioSim™ anti-Evolocumab ELISA kit qualitatively measures antibodies to Evolocumab in human serum or plasma samples. The kit is based on the Sandwich ELISA principle. Standards and samples (serum or plasma) are added to the microtiter plate coated with Evolocumab. After incubation, the wells are washed. The HRP conjugated probe is added and binds to the antibodies to Evolocumab that is captured by Evolocumab coated on the wells. Following incubation, wells are washed and the enzymatic activity is detected by the addition of TMB chromogen substrate. Finally, the reaction is terminated with an acidic stop solution. The color developed is proportional to the amount of antibodies to Evolocumab in the sample or standard. The qualitative test results can be evaluated using cut-off values.

II. Features and Benefits:

- · For in vitro Qualitative determination of antibodies to Evolocumab in human serum and plasma samples
- Assay Precision: Intra-Assay and Inter-Assay CV < 30%
- Cross Reactivity: Except for Evolocumab, there is no cross reaction with other therapeutic antibodies and native serum immunoglobins

III. Sample Type:

Human Plasma and Serum

IV. Kit Contents:

Components	E5027-100	Part Number
Microtiter Plate	8 X 12 Strips	E5027-100-1
Positive Control	0.3 ml	E5027-100-2
Negative Control	1 ml	E5027-100-3
Assay Buffer	12 ml	E5027-100-4
HRP-conjugate Probe	12 ml	E5027-100-5
TMB substrate (Avoid light)	12 ml	E5027-100-6
Stop Solution	12 ml	E5027-100-7
Wash buffer (20X)	50 ml	E5027-100-8
Plate sealers	2	E5027-100-9

V. User Supplied Reagents and Equipment:

- Microplate reader capable of measuring absorbance at 450 nm
- Precision pipettes with disposable tips
- Clean eppendorf tubes for preparing standards and sample dilutions
- Absorbent paper

VI. Storage and Handling:

The entire kit may be stored at 4°C for up to 12 months from the date of shipment

VII. Reagent and Sample Preparation:

Note: Before using the kit, spin the tubes and bring down all the components to the bottom of the tubes

- 1. Wash Buffer: Dilute the Wash Buffer (20X) to 1X solution using ddH₂O (10 ml of Wash Buffer (20X) + 190 ml of ddH₂O). Mix the 1X solution thoroughly by manual vortexing. The working stock is stable for 2 weeks after preparation at 4°C. If there are crystals in Wash Buffer (20X), warm the Buffer at 37°C prior to dilution.
- 2. Sample preparation: The usual precautions should be observed for venipuncture. Samples that are hemolytic, icteric or lipemic should be avoided. If the sample is turbid, then it must be centrifuged to separate particulates from solution. Freeze/thawing of serum/plasma samples should be avoided. Drug infusions may interfere with the detection of antibodies to drugs in serum/plasma samples. Hence, it is advisable to take blood samples prior to the scheduled dose. Collected samples are stable for 2 days at 4°C or for 6 months at -20°C.

VIII. Assay Protocol:

Note: Bring all the reagents, samples and microtiter plate to room temperature 15 minutes prior to the assay. It is recommended that all samples be run at least in duplicates.

05/21



Gentaur Europe BVBA Voortstraat 49, 1910 Kampenhout BELGIUM Tel 0032 16 58 90 45 info@gentaur.com



- 1. Prepare all reagents and samples as instructed in section VII.
- 2. Pipette 100 µl of Assay Buffer into each of the wells to be used.
- 3. Add 10 µl of **Negative control** (2 wells), **Positive control** (1 well), and **samples** into appropriate wells. Cover wells and incubate at room temperature (RT) for 60 minutes.
- 4. Discard the incubation solution. Wash plate 3 times with 300 μl of **Wash Buffer (1X)**. Remove excess solution by tapping the inverted plate on an absorbent paper.
- 5. Add 100 µl of HRP-conjugated probe into each well. Cover the plate with adhesive plate sealer and incubate at RT for 60 minutes.
- 6. Discard the solution and wash the wells as in step 4.
- 7. Add 100 µl of TMB substrate solution and incubate the plate in dark at RT for 20 minutes.
- 8. Add 100 µl of **Stop solution** to stop the reaction. Color changes from blue to yellow.
- 9. Read the absorbance in micro plate reader set to 450 nm within 30 minutes after pipetting the **Stop solution**. (Use reference wavelength as 650 nm)

IX. Interpretation of Results:

For the run to be valid, the OD 450/650 nm of the Positive control should be > 1.500 and the OD 450/650 nm of each Negative control should be < 0.150. If the results do not comply with the aforementioned information, then improper technique or reagent deterioration may be suspected and therefore the assay must be repeated.

The results are evaluated by a cut-off value which is estimated by multiplying the mean OD 450/650 nm of the negative controls by 3

- If "Sample OD450/650 / the mean Negative Control OD450/650" is < 3, the sample is NEGATIVE for antibodies to Evolocumab.
- If "Sample OD450/650 / the mean Negative Control OD450/650" is ≥ 3, the sample is POSITIVE for antibodies to Evolocumab.

Note: The cut-off information provided with this kit can only be considered as a recommendation. Cut-off values must be calculated/set or verified according to scientific standards by the users.

X. Related Products:

- BioSim™ Rituximab (Human) ELISA Kit (Cat. No. E4371-100)
- BioSim™ Adalimumab (Human) ELISA Kit (Cat. No. E4372-100)
- BioSim™ Bevacizumab (Human) ELISA Kit (Cat. No. E4373-100)
- BioSim[™] Etanercept (Human) ELISA Kit (Cat. No. E4374-100)
- BioSim™ Infliximab (Human) ELISA Kit (Cat. No. E4375-100)
- BioSim[™] Trastuzumab (Human) ELISA Kit (Cat. No. E4376-100)
- BioSim™ Golimumab (Human) ELISA Kit (Cat. No. E4377-100)
- BioSim™ Infliximab (Human) ELISA Kit (Cat. No. E4378-100)
- BioSim™ Cetuximab (Human) ELISA Kit (Cat. No. E4379-100)
 BioSim™ Denosumab (Human) ELISA Kit (Cat. No. E4380-100)
- BioSim™ anti-PD-L1 Mab (Human) ELISA Kit (E4922)
- BioSim™ Nivolumab (Human) ELISA Kit (Cat. No. E4382-100)
- BioSim™ Pembrolizumab (Human) ELISA Kit (Cat. No. E4383-100)
- BioSim[™] Avelumab (Human) ELISA Kit (Cat. No. E4384-100)
- BioSim™ Durvalumab (Human) ELISA Kit (E4925)
- BioSim™ anti-Durvalumab (Human) ELISA Kit (E4926)
- BioSim™ anti-Pembrolizumab (Human) ELISA Kit (E4397)
- BioSim™ anti-Nivolumab (Human) ELISA Kit (E4396)