

# Caspase-2 Flu

(Catalog #K116-25, -100, -200, -400; Store kit at -20 °C)

## I. Introduction:

Activation of ICE-family proteases/caspases initiates apoptosis in mammalian cells. The **Caspase-2 Fluorometric Assay Kit** provides a simple and convenient means for assaying the activity of caspases that recognize the sequence VDVAD. The assay is based on detection of cleavage of substrate VDVAD-AFC (AFC: 7-amino-4-trifluoromethyl coumarin). VDVAD-AFC emits blue light ( $\lambda_{max} = 400 \text{ nm}$ ); upon cleavage of the substrate by caspase-2 or related caspases, free AFC emits a yellow-green fluorescence ( $\lambda_{max} = 505 \text{ nm}$ ), which can be quantified using a fluorometer or fluorescence microtiter plate reader. Comparison of the fluorescence of AFC from an apoptotic sample with an uninduced control allows determination of the fold increase in Caspase-2 activity.

## II. Kit Contents:

Components	K116-25	K116-100	K116-200	K116-400	Part Number
	25 assays	100 assays	200 assays	400 assays	
Cell Lysis Buffer	25 ml	100 ml	100 ml	100 ml	K116-XX(X)-1
2X Reaction Buffer	2 ml	4 x 2 ml	16 ml	32 ml	K116-XX(X)-2
VDVAD-AFC (1 mM)	125 $\mu\text{l}$	0.5 ml	2 x 0.5 ml	2 x 1 ml	K116-XX(X)-3
DTT (1 M)	100 $\mu\text{l}$	0.4 ml	0.4 ml	0.4 ml	K116-XX(X)-4

## III. Caspase-2 Assay Protocol:

### A. General Considerations

- Aliquot enough 2X Reaction Buffer for the number of assays to be performed. Add DTT to the 2X Reaction Buffer immediately before use (10 mM final concentration: add 10  $\mu\text{l}$  of 1.0 M DTT stock per 1 ml of 2X Reaction Buffer).
- After thawing, store the Cell Lysis Buffer and 2X Reaction Buffer at 4 °C. All kit reagents are stable for 6 months
- Protect VDVAD-AFC from light.
- We recommend using a flat bottom, opaque, white or black 96-well plate for enhanced sensitivity.

### B. Assay Procedure:

1. Induce apoptosis in cells by desired method. Concurrently incubate a control culture *without* induction.
2. Count cells and pellet 1-5 x 10<sup>6</sup> cells or use 50-200  $\mu\text{g}$  cell lysates if protein concentration has been measured.
3. Resuspend cells in 50  $\mu\text{l}$  of chilled Cell Lysis Buffer. Incubate cells on ice for 10 min.
4. Centrifuge at 10000 g for 3-5 min at 4 °C. Use the cleared cell lysate for the assay.
5. Add 50  $\mu\text{l}$  of 2X Reaction Buffer (containing 10 mM DTT) to each sample. Add 5  $\mu\text{l}$  of the 1 mM VDVAD-AFC substrate (50  $\mu\text{M}$  final concentration) and incubate at 37 °C for 1-2 hr.
6. Read samples in a fluorometer equipped with a 400 nm excitation filter and 505 nm emission filter. For a plate-reading set-up, transfer the samples to a 96-well plate. You may also perform the entire assay directly in a 96-well plate. Fold-increase in Caspase-2 activity can be determined by comparing the results of treated samples with the level of the uninduced control.

## RELATED PRODUCTS:

### Apoptosis Detection Kits & Reagents

- Annexin V Kits & Bulk Reagents
- Caspase Assay Kits & Reagents
- Mitochondrial Apoptosis Kits & Reagents
- Nuclear Apoptosis Kits & Reagents
- Apoptosis Inducers and Set
- Apoptosis siRNA Vectors

### Cell Fractionation System

- Mitochondria/Cytosol Fractionation Kit
- Nuclear/Cytosol Fractionation Kit
- Membrane Protein Extraction Kit
- Cytosol/Particulate Rapid Separation Kit
- Mammalian Cell Extraction Kit
- FractionPREP Fractionation System

### Cell Proliferation & Senescence

- Quick Cell Proliferation Assay Kit
- Senescence Detection Kit
- High Throughput Apoptosis/Cell Viability Assay Kits
- LDH-Cytotoxicity Assay Kit
- Bioluminescence Cytotoxicity Assay Kit
- Live/Dead Cell Staining Kit

### Cell Damage & Repair

- HDAC Fluorometric & Colorimetric Assays & Drug Discovery Kits
- HAT Colorimetric Assay Kit & Reagents
- DNA Damage Quantification Kit
- Glutathione & Nitric Oxide Fluorometric & Colorimetric Assay Kits

### Signal Transduction

- cAMP & cGMP Assay Kits
- Akt & JNK Activity Assay Kits
- Beta-Secretase Activity Assay Kit

### Adipocyte & Lipid Transfer

- Recombinant Adiponectin, Survivin, & Leptin
- CETP Activity Assay & Drug Discovery Kits
- PLTP Activity Assay & Drug Discovery Kits
- Total Cholesterol Quantification Kit

### Molecular Biology & Reporter Assays

- siRNA Expression Vectors
- Cloning Insert Quick Screening Kit
- Mitochondrial & Genomic DNA Isolation Kits
- 5 Minutes DNA Ligation Kit
- 20 Minutes Gel Staining/Destaining Kit
- $\beta$ -Galactosidase Staining Kit & Luciferase Reporter Assay Kit

### Growth Factors and Cytokines

**FOR RESEARCH USE ONLY! Not to be used on humans.**

Problems	Cause	Solution
Assay not working	<ul style="list-style-type: none"> <li>• Cells did not lyse completely</li> <li>• Experiment was not performed at optimal time after apoptosis induction</li> <li>• Plate read at incorrect wavelength</li> <li>• Old DTT used</li> </ul>	<ul style="list-style-type: none"> <li>• Resuspend the cell pellet in the lysis buffer and incubate as described in the datasheet</li> <li>• Perform a time-course induction experiment for apoptosis</li> <li>• Check the wavelength listed in the datasheet and the filter settings of the instrument</li> <li>• Always use freshly thawed DTT in the cell lysis buffer</li> </ul>
High Background	<ul style="list-style-type: none"> <li>• Increased amount of cell lysate used</li> <li>• Increased amounts of components added due to incorrect pipetting</li> <li>• Incubation of cell samples for extended periods</li> <li>• Use of expired kit or improperly stored reagents</li> <li>• Contaminated cells</li> </ul>	<ul style="list-style-type: none"> <li>• Refer to datasheet and use the suggested cell number to prepare lysates</li> <li>• Use calibrated pipettes</li> <li>• Refer to datasheet and incubate for exact times</li> <li>• Always check the expiry date and store the individual components appropriately</li> <li>• Check for bacterial/ yeast/ mycoplasma contamination</li> </ul>
Lower signal levels	<ul style="list-style-type: none"> <li>• Cells did not initiate apoptosis</li> <li>• Very few cells used for analysis</li> <li>• Use of samples stored for a long time</li> <li>• Incorrect setting of the equipment used to read samples</li> <li>• Allowing the reagents to sit for extended times on ice</li> </ul>	<ul style="list-style-type: none"> <li>• Determine the time-point for initiation of apoptosis after induction (time-course experiment)</li> <li>• Refer to datasheet for appropriate cell number</li> <li>• Use fresh samples or aliquot and store and use within one month for the assay</li> <li>• Refer to datasheet and use the recommended filter setting</li> <li>• Always thaw and prepare fresh reaction mix before use</li> </ul>
Samples with erratic readings	<ul style="list-style-type: none"> <li>• Uneven number of cells seeded in the wells</li> <li>• Samples prepared in a different buffer</li> <li>• Adherent cells dislodged and lost at the time of experiment</li> <li>• Cell/ tissue samples were not completely homogenized</li> <li>• Samples used after multiple freeze-thaw cycles</li> <li>• Presence of interfering substance in the sample</li> <li>• Use of old or inappropriately stored samples</li> </ul>	<ul style="list-style-type: none"> <li>• Seed only equal number of healthy cells (correct passage number)</li> <li>• Use the cell lysis buffer provided in the kit</li> <li>• Perform experiment gently and in duplicates/triplicates; apoptotic cells may become floaters</li> <li>• Use Dounce homogenizer (increase the number of strokes); observe efficiency of lysis under microscope</li> <li>• Aliquot and freeze samples, if needed to use multiple times</li> <li>• Troubleshoot as needed</li> <li>• Use fresh samples or store at correct temperatures until use</li> </ul>
Unanticipated results	<ul style="list-style-type: none"> <li>• Measured at incorrect wavelength</li> <li>• Cell samples contain interfering substances</li> </ul>	<ul style="list-style-type: none"> <li>• Check the equipment and the filter setting</li> <li>• Troubleshoot if it interferes with the kit (run proper controls)</li> </ul>
General issues	<ul style="list-style-type: none"> <li>• Improperly thawed components</li> <li>• Incorrect incubation times or temperatures</li> <li>• Incorrect volumes used</li> <li>• Air bubbles formed in the well/tube</li> <li>• Substituting reagents from older kits/ lots</li> <li>• Use of a different 96-well plate</li> </ul>	<ul style="list-style-type: none"> <li>• Thaw all components completely and mix gently before use</li> <li>• Refer to datasheet &amp; verify the correct incubation times and temperatures</li> <li>• Use calibrated pipettes and aliquot correctly</li> <li>• Pipette gently against the wall of the well/tubes</li> <li>• Use fresh components from the same kit</li> <li>• Fluorescence: Black plates; Absorbance: Clear plates</li> </ul>

**Note:** The most probable cause is listed under each section. Causes may overlap with other sections.