

Caspase-Family Fluorometric Substrate Set Plus

CATALOG #: K135-7-25

LOT #: _____

STORAGE CONDITIONS: Store at -20°C.

SHELF LIFE: 6 months under proper storage conditions

KIT CONTENTS:

Concentration	Description	Volume	Part Number
1 mM	Caspase-1 Substrate, Ac-YVAD-AFC	125 µl	K110-25-3
1 mM	Caspase-2 Substrate, Ac-VDVAD-AFC	125 µl	K116-25-3
1 mM	Caspase-3 Substrate, Ac-DEVD-AFC	125 µl	K105-25-3
1 mM	Caspase-5 Substrate, Ac-WEHD-AFC	125 µl	K122-25-3
1 mM	Caspase-6 Substrate, Ac-VEID-AFC	125 µl	K114-25-3
1 mM	Caspase-8 Substrate, Ac-IETD-AFC	125 µl	K112-25-3
1 mM	Caspase-9 Substrate, Ac-LEHD-AFC	125 µl	K118-25-3
N/A	Cell Lysis Buffer	100 ml	1067-100
N/A	2X Reaction Buffer	20 ml	1068-20
1 M	DTT	0.4 ml	K110-100-4

ASSAY PROCEDURE:

1. Induce apoptosis in cells by desired method. Concurrently incubate a control culture *without* induction.
2. Count cells and pellet 1-5 x 10⁶ cells or use 50-200µg cell lysates if protein concentration has been measured.
3. Resuspend cells in 50 µl of chilled Cell Lysis Buffer.
4. Incubate cells on ice for 10 minutes.
5. Add 50 µl of 2X Reaction Buffer and 1 µl DTT to each sample.
6. Add 5 µl of the 1 mM AFC conjugated substrates (50 µM final conc.) into each tube individually and incubate at 37°C for 1-2 hour.
7. Read samples in a fluorometer equipped with a 400-nm excitation filter and 505-nm emission filter. For a plate-reading set-up, transfer the samples to a 96-well plate. You may perform the entire assay directly in a 96-well plate.
Fold-increase in caspase activity can be determined by comparing these results with the level of the uninduced control.

RELATED PRODUCTS:

Apoptosis Detection Kits & Reagents

- Annexin V Kits & Bulk Reagents
- Caspase Assay Kits & Reagents
- Mitochondrial Apoptosis Kits & Reagents
- Nuclear Apoptosis Kits & Reagents
- Apoptosis Inducers and Set
- Apoptosis siRNA Vectors

Cell Fractionation System

- Mitochondria/Cytosol Fractionation Kit
- Nuclear/Cytosol Fractionation Kit
- Membrane Protein Extraction Kit
- Cytosol/Particulate Rapid Separation Kit
- Mammalian Cell Extraction Kit
- FractionPREP Fractionation System

Cell Proliferation & Senescence

- Quick Cell Proliferation Assay Kit
- Senescence Detection Kit
- High Throughput Apoptosis/Cell Viability Assay Kits
- LDH-Cytotoxicity Assay Kit
- Bioluminescence Cytotoxicity Assay Kit
- Live/Dead Cell Staining Kit

Cell Damage & Repair

- HDAC Fluorometric & Colorimetric Assays & Drug Discovery Kits
- HAT Colorimetric Assay Kit & Reagents
- DNA Damage Quantification Kit
- Glutathione & Nitric Oxide Fluorometric & Colorimetric Assay Kits

Signal Transduction

- cAMP & cGMP Assay Kits
- Akt & JNK Activity Assay Kits
- Beta-Secretase Activity Assay Kit

Adipocyte & Lipid Transfer

- Recombinant Adiponectin, Survivin, & Leptin
- CETP Activity Assay & Drug Discovery Kits
- PLTP Activity Assay & Drug Discovery Kits
- Total Cholesterol Quantification Kit

Molecular Biology & Reporter Assays

- siRNA Vectors
- Cloning Insert Quick Screening Kit
- Mitochondrial & Genomic DNA Isolation Kits
- 5 Minutes DNA Ligation Kit
- 20 Minutes Gel Staining/Destaining Kit
- β -Galactosidase Staining Kit & Luciferase Reporter Assay Kit

Growth Factors and Cytokines

Monoclonal and Polyclonal Antibodies

GENERAL TROUBLESHOOTING GUIDE FOR CASPASE COLORIMETRIC AND FLUORESCENT KITS

Problems	Cause	Solution
Assay not working	<ul style="list-style-type: none"> • Cells did not lyse completely • Experiment was not performed at optimal time after apoptosis induction • Plate read at incorrect wavelength • Old DTT used 	<ul style="list-style-type: none"> • Resuspend the cell pellet in the lysis buffer and incubate as described in the datasheet • Perform a time-course induction experiment for apoptosis • Check the wavelength listed in the datasheet and the filter settings of the instrument • Always use freshly thawed DTT in the cell lysis buffer
High Background	<ul style="list-style-type: none"> • Increased amount of cell lysate used • Increased amounts of components added due to incorrect pipetting • Incubation of cell samples for extended periods • Use of expired kit or improperly stored reagents • Contaminated cells 	<ul style="list-style-type: none"> • Refer to datasheet and use the suggested cell number to prepare lysates • Use calibrated pipettes • Refer to datasheet and incubate for exact times • Always check the expiry date and store the individual components appropriately • Check for bacteria/ yeast/ mycoplasma contamination
Lower signal levels	<ul style="list-style-type: none"> • Cells did not initiate apoptosis • Very few cells used for analysis • Use of samples stored for a long time • Incorrect setting of the equipment used to read samples • Allowing the reagents to sit for extended times on ice 	<ul style="list-style-type: none"> • Determine the time-point for initiation of apoptosis after induction (time-course experiment) • Refer to datasheet for appropriate cell number • Use fresh samples or aliquot and store and use within one month for the assay • Refer to datasheet and use the recommended filter setting • Always thaw and prepare fresh reaction mix before use
Samples with erratic readings	<ul style="list-style-type: none"> • Uneven number of cells seeded in the wells • Samples prepared in a different buffer • Adherent cells dislodged and lost at the time of experiment • Cell/ tissue samples were not completely homogenized • Samples used after multiple freeze-thaw cycles • Presence of interfering substance in the sample • Use of old or inappropriately stored samples 	<ul style="list-style-type: none"> • Seed only equal number of healthy cells (correct passage number) • Use the cell lysis buffer provided in the kit • Perform experiment gently and in duplicates/triplicates; apoptotic cells may become floaters • Use Dounce homogenizer (increase the number of strokes); observe efficiency of lysis under microscope • Aliquot and freeze samples, if needed to use multiple times • Troubleshoot as needed • Use fresh samples or store at correct temperatures until use
Unanticipated results	<ul style="list-style-type: none"> • Measured at incorrect wavelength • Cell samples contain interfering substances 	<ul style="list-style-type: none"> • Check the equipment and the filter setting • Troubleshoot if it interferes with the kit (run proper controls)
General issues	<ul style="list-style-type: none"> • Improperly thawed components • Incorrect incubation times or temperatures • Incorrect volumes used • Air bubbles formed in the well/tube • Substituting reagents from older kits/ lots • Use of a different 96-well plate 	<ul style="list-style-type: none"> • Thaw all components completely and mix gently before use • Refer to datasheet & verify the correct incubation times and temperatures • Use calibrated pipettes and aliquot correctly • Pipette gently against the wall of the well/tubes • Use fresh components from the same kit • Fluorescence: Black plates; Absorbance: Clear plates

Note# The most probable cause is listed under each section. Causes may overlap with other sections.