



# Caspase-6 Inhibitor Drug Screening Kit (Fluorometric)

(Catalog #K156-100; 100 assays; Store kit at -20 °C)

#### I. Introduction:

Caspases have been shown to play a crucial role in apoptosis induced by various deleterious and physiologic stimuli. Inhibition of caspases can delay apoptosis, implicating a potential role in drug screening efforts. The Caspase-6 Inhibitor Drug Screening Kit provides an effective means for screening caspase inhibitors using fluorometric methods. The assay utilizes synthetic peptide substrate VEID-AFC (AFC, 7-amino-4-trifluoromethyl coumarin). Active caspase-6 cleaves the synthetic substrate to release free AFC which can then be quantified by fluorometry. Compounds to be screened can directly be added to the reaction and the level of inhibition of caspase-6 activity can be determined by comparison of the fluorescence intensity in samples with and without the testing inhibitors. The assay is simple, straightforward, and can be performed directly in microtiter plates. Each kit contains 100 units of active caspase-6, sufficient for screening 100 caspase inhibitor samples. Assay conditions have been optimized to obtain the maximal activity.

#### II. Kit Contents:

| Components                          | K156-100  | Cap Code    | Part Number |
|-------------------------------------|-----------|-------------|-------------|
| 2X Reaction Buffer                  | 10 ml     | NM          | K156-100-1  |
| Caspase Substrate VEID-AFC (1 mM)   | 0.5 ml    | Amber       | K156-100-2  |
| DTT (1 M)                           | 100 µl    | Green       | K156-100-3  |
| Active Caspase-6 (Lyophilized)      | 100 units | Purple      | K156-100-4  |
| Caspase Inhibitor, Z-VAD-FMK (2 mM) | 10 µl     | Clear/Amber | K156-100-5  |

#### III. Caspase-6 Assay Protocol:

#### A. General Considerations & Reagent Preparations

- After thawing, store the 2X Reaction Buffer at 4 °C. Aliquot enough 2X Reaction Buffer for the number of assays to be performed. Add DTT to the 2X Reaction Buffer immediately before use (10 mM final concentration: add 10 µl of 1.0 M DTT stock per 1 ml of 2X Reaction Buffer).
- Protect VEID-AFC from light.
- Reconstitute the Active Caspase-6 in 550 µl 2X Reaction Buffer. Aliquot and immediately store at 70 °C.
- We recommend using a flat bottom, opaque, white or black 96-well plate for enhanced sensitivity.

### B. Assay Procedure

1. Prepare testing sample in  $dH_2O$  to a final volume of 50  $\mu$ I/well. Add 5  $\mu$ I of Active Caspase-6. Mix well.

Prepare a background control by omitting the Active Caspase-6 from the reaction mixture. Prepare a positive inhibition control by adding 1 µI of the Caspase-6 Inhibitor (provided with the kit) instead of your testing inhibitor.

2. Prepare a Master Mix for each assay containing the follows:

45 µl 2X Reaction Buffer (containing 10 mM DTT)

5 μl 1 mM VEID-AFC substrate (50 μM final concentration)

- 3. Mix well and add 50  $\mu$ l of the Master Mix to each well to start the reaction.
- 4. Incubate at 37 °C for 0.5-1 hour.
- Read samples in a fluorescence plate reader equipped with a 400 nm excitation filter and 505 nm emission filters. Comparison of the fluorescence intensity of the testing samples with samples containing no inhibitors to determine the inhibition efficiency of the testing inhibitors.

### IV. Storage and Stability:

Store the kit at -20 °C. Store 2X Reaction Buffer at 4 °C after opening. All reagents are stable for one year under proper storage conditions.

#### **RELATED PRODUCTS:**

Gentaur Europe BVBA Voortstraat 49, 1910 Kampenhout BELGIUM Tel 0032 16 58 90 45 info@gentaur.com

Apoptosis Detection Kits & Reagents

- Annexin V Kits & Bulk Reagents
- Caspase Assay Kits & Reagents
- Mitochondrial Apoptosis Kits & Reagents
- Nuclear Apoptosis Kits & Reagents
- Additional Apoptosis Kits & Reagents

#### Cell Fractionation System

- Mitochondria/Cytosol Fractionation Kit
- Nuclear/Cytosol Fractionation Kit
- Membrane Protein Extraction Kit
- Cytosol/Particulate Rapid Separation Kit
- Mammalian Cell Extraction Kit
- FractionPREP Fractionation System

### Cell Proliferation & Senescence

- Quick Cell Proliferation Assay Kit
- Senescence Detection Kit
- High Throughput Apoptosis/Cell Viability Assay Kits
- LDH-Cytotoxicity Assay Kit
- Bioluminescence Cytotoxicity Assay Kit
- Live/Dead Cell Staining Kit

#### Cell Damage & Repair

- HDAC Fluorometric & Colorimetric Assays & Drug Discovery Kits
- HAT Colorimetric Assay Kit & Reagents
- DNA Damage Quantification Kit
- Glutathione Fluorometric & Colorimetric Assay Kits
- Nitric Oxide Fluorometric & Colorimetric Assay Kits

### Signal Transduction

- Camp & cGMP Assay Kits
- Akt & JNK Activity Assay Kits
- Beta-Secretase Activity Assay Kit

### Adipocyte & Lipid Transfer

- Recombinant Adiponectin, Survivin, & Leptin
- CETP Activity Assay & Drug Discovery Kits
- Total Cholesterol Quantification Kit

### Molecular Biology & Reporter Assavs

- siRNA Vectors
- Cloning Insert Quick Screening Kit
- Mitochondrial & Genomic DNA Isolation Kits

#### FOR RESEARCH USE ONLY! Not to be used on humans.

Page 1





#### Gentaur Europe BVBA Voortstraat 49, 1910 Kampenhout BELGIUM Tel 0032 16 58 90 45 info@gentaur.com

## GENERAL TROUBLESHOOTING GUIDE FOR CASPASE INHIBITOR DRUG SCREENING KITS:

| Problems                         | Cause  | Solution   |  |
|----------------------------------|--|--|--|
| Assay not working                | Inactive Caspases due to incorrect reconstitution and storage        | Reconstitute in reaction buffer, aliquot and store as described in the datasheet |  |
|                                  | Use of degraded Caspase substrate                                    | Protect tube from direct light and store appropriately                           |  |
|                                  | Plate read at incorrect wavelength                                   | Check the wavelength listed in the datasheet and the filter settings of the      |  |
|                                  | Old DTT used   | instrument Always use freshly thawed DTT   |  |
| High Background                  | Increased amounts of components added due to incorrect pipetting     | Use calibrated pipettes  |  |
|                                  | Use of substrate that has been exposed to light for extended periods | Store and handle substrate as indicated in the data sheet Page 2                 |  |
| Lower signal levels              | Incorrect setting of the equipment used to read samples              | Refer to datasheet and use the recommended filter setting                        |  |
|                                  | Allowing the reagents to sit for extended times on ice               | Always thaw and prepare fresh reaction mix before use                            |  |
| Samples with erratic readings    | Drugs tested at lower/ higher concentrations                         | Refer literature and use appropriate concentrations; test several concentrations |  |
|                                  | Drugs prepared in a different buffer                                 | Check if the components of the buffer could inhibit the reaction                 |  |
|                                  | Presence of interfering substance in the drug sample                 | Troubleshoot as needed   |  |
|                                  | Measured at incorrect wavelength                                     | Check the equipment and the filter setting                                       |  |
|                                  | Drug samples contains interfering substances                         | Troubleshoot if it interferes with the kit (run proper controls)                 |  |
| General issues                   | Improperly thawed components   | Thaw all components completely and mix gently before use                         |  |
|                                  | Incorrect incubation times or temperatures                           | • Refer to datasheet & verify the correct incubation times and temperatures      |  |
|                                  | Incorrect volumes used   | Use calibrated pipettes and aliquot correctly                                    |  |
|                                  | Air bubbles formed in the well/tube                                  | Pipette gently against the wall of the well/tubes                                |  |
|                                  | Substituting reagents from older kits/ lots                          | Use fresh components from the same kit   |  |
|                                  | Use of a different 96-well plate                                     | Fluorescence: Black plates; Absorbance: Clear plates                             |  |
| Note: The most probable cause is | listed under each section. Causes may overlap with other sections.   |  |  |