

# CaspGLOW™ Fluorescein Active Caspase-2 Staining Kit

(Catalog# K182-25, -100; Store kit at -20° C)

## I. Introduction:

Activation of caspases plays a central role in apoptosis. The **CaspGLOW™ Fluorescein Active Caspase-2 Staining Kit** provides a convenient means for sensitive detection of activated caspase-2 in living cells. The assay utilizes the caspase-2 inhibitor, VDVAD-FMK, conjugated to FITC (FITC-VDVAD-FMK) as a marker. FITC-VDVAD-FMK is cell permeable, nontoxic, and irreversibly binds to activated caspase-2 in apoptotic cells. The FITC label allows for direct detection of activated caspases in apoptotic cells by fluorescence microscopy, flow cytometry, or fluorescence plate reader.

## II. Kit Contents:

Components	K182-25	K182-100	Part Number
	25 assays	100 assays	
FITC-VDVAD-FMK	25 µl	100 µl	K182-xx(x)-1
Wash Buffer	50 ml	2 x 100 ml	K182-xx(x)-2
Z-VAD-FMK	10 µl	10 µl	K182-xx(x)-3

## III. Caspase-2 Assay Procedure:

### A. Staining Procedure:

1. Induce apoptosis in cells (1 x 10<sup>6</sup>/ml) by desired method. Concurrently incubate a control culture *without* induction. An additional negative control can be prepared by adding the caspase inhibitor Z-VAD-FMK at 1 µl/ml to an induced culture to inhibit caspase-2 activation.
2. Aliquot 300 µl each of the induced and control cultures into eppendorf tubes.
3. Add 1 µl of FITC-VDVAD-FMK into each tube and incubate for 0.5-1 hour at 37° C incubator with 5% CO<sub>2</sub>.
4. Centrifuge cells at 3000 rpm for 5 minutes and remove supernatant.
5. Resuspend cells in 0.5 ml of Wash Buffer, and centrifuge again.
6. Repeat Step 5.  
Proceed to B, C, or D depending on methods of analysis.

### B. Quantification by Flow Cytometry:

For flow cytometric analysis, resuspend cells in 300 µl of Wash Buffer. Keep samples on ice. Analyzing samples by flow cytometry using the FL-1 channel.

### C. Detection by Fluorescence Microscopy:

For fluorescence microscopic analysis, resuspend cells in 100 µl Wash Buffer. Transfer one drop of the cell suspension onto a microslide and cover with a coverslip. Observe cells under a fluorescence microscope using FITC filter. Caspase positive cells appear to have brighter green signals, whereas caspase negative control cells show much weaker signal.

## D. Analysis by Fluorescence Plate Reader:

For analysis with fluorescence plate reader, resuspend cells in 100 µl Wash Buffer and then transfer the cell suspension into each well in the black microtiter plate. Measure the fluorescence intensity at Ex. = 485 nm and Em. = 535 nm. For control, use wells containing unlabeled cells.

## RELATED PRODUCTS:

- CaspGLOW Fluorescein Active Caspase Staining Kit (Cat. No. K180-25, 100)
- CaspGLOW Fluorescein Active Caspase-12 Staining Kit (Cat. No. K172-25, 100)
- CaspGLOW Fluorescein Active Caspase-3 Staining Kit (Cat. No. K183-25, 100)
- CaspGLOW Fluorescein Active Caspase-8 Staining Kit (Cat. No. K188-25, 100)
- CaspGLOW Fluorescein Active Caspase-9 Staining Kit (Cat. No. K189-25, 100)
- CaspGLOW Red Active Caspase Staining Kit (Cat. No. K190-25, 100)
- CaspGLOW Red Active Caspase-3 Staining Kit (Cat. No. K193-25, 100)
- CaspGLOW Red Active Caspase-8 Staining Kit (Cat. No. K198-25, 100)
- CaspGLOW Red Active Caspase-9 Staining Kit (Cat. No. K199-25, 100)

**FOR RESEARCH USE ONLY! Not to be used on humans.**



**GENERAL TROUBLESHOOTING GUIDE FOR CaspGLOW BASED ASSAYS:**

Problems	Cause	Solution
<b>High background</b>	<ul style="list-style-type: none"> <li>• Cell density is higher than recommended</li> <li>• Cells were not washed well with wash buffer after staining</li> <li>• Cells were Incubated for extended period of time</li> <li>• Use of extremely confluent cells</li> <li>• Cells were contaminated</li> </ul>	<ul style="list-style-type: none"> <li>• Refer to datasheet and use the suggested cell number</li> <li>• Use the wash buffer provided, and as instructed in the datasheet</li> <li>• Refer to datasheets for proper incubation time</li> <li>• Perform assay when cells are at 70-95% confluency</li> <li>• Check for bacteria/ yeast/ mycoplasma contamination</li> </ul>
<b>Lower signal level</b>	<ul style="list-style-type: none"> <li>• Cells did not initiate apoptosis</li> <li>• Very few cells were used for analysis</li> <li>• Incorrect setting of the equipment or wavelength used to read samples</li> <li>• Use of expired kit or improperly stored reagents</li> </ul>	<ul style="list-style-type: none"> <li>• Determine the optimal time and dose for apoptosis induction (time-course experiment)</li> <li>• Refer to data sheet for appropriate cell number</li> <li>• Refer to datasheet and use the recommended filter setting</li> <li>• Always check the expiry date and store the components appropriately</li> </ul>
<b>Erratic results</b>	<ul style="list-style-type: none"> <li>• Old (unhealthy) cells used</li> <li>• Adherent cells were dislodged and washed away prior to assaying</li> <li>• Incorrect incubation times or temperatures</li> <li>• Incorrect volumes used</li> </ul>	<ul style="list-style-type: none"> <li>• Seed healthy cells and make sure cells are healthy prior to induction of apoptosis</li> <li>• Collect all cells (both attached and dislodged) after induction for accurate results</li> <li>• Refer to datasheet &amp; verify correct incubation times and temperatures</li> <li>• Use calibrated pipettes and aliquot correctly</li> </ul>
<p><b>Note:</b> The most probable cause is listed under each section. Causes may overlap with other sections.</p>		