




Succinyl-CoA Synthetase Activity Colorimetric Assay Kit

9/14

(Catalog # K597-100; 100 assays; Store at -20°C)

I. Introduction:

Succinyl-CoA Synthetase (SCS, also called Succinyl-CoA ligase, Succinate Thiokinase) (EC 6.2.1.5) is a critical enzyme in the citric acid cycle and an important metabolic intermediate for porphyrin, heme and ketone body biosynthesis. It is located in the mitochondrial matrix and is a heterodimer composed of one α and one β subunit. In humans, Succinyl-CoA Synthetase deficiency causes the build-up of lactic acid leading to lactic acidosis, which can be fatal in infants. Measurement and analysis of SCS activity is useful for both mechanistic studies as well as for diagnostic purposes. In BioVision's Succinyl-CoA Synthetase Activity Assay, SCS converts succinate into succinyl-CoA in the presence of ATP and CoA. Succinyl-CoA reacts with the Developer to form a colored product with strong absorbance at 450 nm. This assay kit is simple, sensitive, and high-throughput adaptable. It can detect less than 0.1 mU of Succinyl-CoA Synthetase activity in a variety of samples.



II. Application:

- Measurement of Succinyl-CoA Synthetase activity in various tissues/cells
- Analysis of cell signaling pathway

III. Sample Type:

- Animal tissues: heart, liver, muscle, etc.
- Purified mitochondria
- Cell culture: adherent or suspension cells

IV. Kit Contents:

Components	K597-100	Cap Code	Part Number
SCS Assay Buffer	25 ml	WM	K597-100-1
SCS Substrate Mix (Lyophilized)	1 vial	Blue	K597-100-2
SCS Enzyme Mix (Lyophilized)	1 vial	Green	K597-100-3
SCS Developer (Lyophilized)	1 vial	Red	K597-100-4
NADH Standard (Lyophilized)	1 vial	Yellow	K597-100-5
SCS Positive Control (Lyophilized)	1 vial	Orange	K597-100-6

V. User Supplied Reagents and Equipment:

- 96-well clear plate with flat bottom
- Multi-well spectrophotometer (ELISA reader)

VI. Storage Conditions and Reagent Preparation:

Store kit at -20°C, protected from light. Briefly centrifuge small vials prior to opening. Read the entire protocol before performing the assay.

- **SCS Assay Buffer:** Warm to room temperature before use. Store at either 4°C or -20°C.
- **SCS Substrate Mix and SCS Developer:** Reconstitute with 220 μ l dH₂O. Pipette up and down to dissolve completely. Store at -20°C. Keep on ice while in use. Use within two months.
- **SCS Enzyme Mix:** Reconstitute with 220 μ l SCS Assay Buffer. Gently pipette up and down to dissolve completely. Store at -20°C. Use within two months.
- **NADH Standard:** Reconstitute with 400 μ l dH₂O to generate 1.25 mM NADH Standard solution. Aliquot and store at -20°C. Keep on ice while in use. Use within two months.
- **SCS Positive Control:** Reconstitute with 100 μ l SCS Assay Buffer and mix thoroughly. Aliquot and store at -70°C. Keep on ice while in use. Use within two months.

VII. Succinyl-CoA Synthetase Activity Assay Protocol:

1. Sample Preparation: Rapidly homogenize tissue (10 mg) or cells (1 x 10⁶) with 100 μ l ice cold SCS Assay Buffer, and keep on ice for 10 min. Centrifuge at 10,000 x g for 5 min. and transfer the supernatant to a fresh tube. Add 5-50 μ l sample per well & adjust the volume to 50 μ l with SCS Assay Buffer. To check SCS activity in mitochondria, isolate the mitochondria from fresh tissue or cells using BioVision's Mitochondria Isolation Kit for Tissue and Cultured Cells (K288). Add 5-50 μ l of isolated mitochondria per well and adjust the volume to 50 μ l with SCS Assay Buffer. For the SCS Positive Control, add 1-10 μ l of SCS Positive Control into desired well(s) and adjust the volume to 50 μ l with SCS Assay Buffer.

Note:

- For unknown samples, we suggest doing a pilot experiment & testing several doses to ensure the readings are within the Standard Curve range.
 - For samples exhibiting background, prepare parallel sample well(s) as background control.
- 2. NADH Standard Curve:** Add 0, 2, 4, 6, 8 and 10 μ l of 1.25 mM NADH Standard into a series of wells in 96 well plate to generate 0, 2.5, 5.0, 7.5, 10 and 12.5 nmol/well of NADH Standard. Adjust the volume to 50 μ l per well with SCS Assay Buffer.

