




# PicoProbe™ Triglyceride (TG) Fluorometric Assay Kit

rev. 1/19

(Catalog # K614-100; 100 assays; Store at -20°C)

## I. Introduction:

Triglycerides (TG) are the main constituents of vegetable oil, animal fat, LDL and VLDL, and play an important role as transporters of fatty acids as well as an energy source. TG is broken down into fatty acids and glycerol, after which both can serve as substrates for energy producing and metabolic pathways. High blood levels of TG are implicated in atherosclerosis, heart disease and stroke as well as in pancreatitis. Biovision's PicoProbe™ Triglyceride Assay kit is suitable for measuring triglyceride levels in samples, which contain reducing substances that may interfere with oxidase-based assays. In this assay, TG is hydrolyzed to glycerol and fatty acid. The glycerol reacts with Triglyceride Enzyme Mix to form an intermediate product, which in turn reacts with PicoProbe & Developer to generate the fluorescence. The generated fluorescence is directly proportional to the amount of triglycerides. This high-throughput adaptable assay kit is simple, sensitive and easy to use & can detect less than 0.4 μM triglycerides in a variety of samples.



## II. Application:

- Measurement of triglycerides in various tissues, cells & other biological fluids
- Analysis of lipid metabolism
- Mechanistic study of cardiovascular disease

## III. Sample Type:

- Animal tissues: Liver, pancreas, heart etc.
- Cell culture: adipocytes etc.
- Biological fluids: serum, plasma & saliva.

## IV. Kit Contents:

Components	K614-100	Cap Code	Part Number
Triglyceride Assay Buffer	25 ml	WM	K614-100-1
PicoProbe™ (in DMSO)	0.4 ml	Blue/Amber	K614-100-2
Lipase (Lyophilized)	1 vial	Blue	K614-100-3
Triglyceride Enzyme Mix (Lyophilized)	1 vial	Green	K614-100-4
Triglyceride Developer (Lyophilized)	1 vial	Red	K614-100-5
Triglyceride Standard (1mM)	0.3 ml	Yellow	K614-100-6

## V. User Supplied Reagents and Equipment:

- 96-well white plate with flat bottom
- Multi-well spectrophotometer (ELISA reader)

## VI. Storage and Handling:

Store kit at -20°C, protected from light. Warm Assay Buffer to room temperature before use. Briefly centrifuge small vials prior to opening. Read the entire protocol before starting the experiment.

## VII. Reagent Preparation and Storage Conditions:

- **PicoProbe™**: Ready to use as supplied. Warm to room temperature before use. Store at -20°C.
- **Lipase and Triglyceride Enzyme Mix**: Reconstitute with 220 μl Triglyceride Assay Buffer. Pipette up and down to dissolve completely. Aliquot and store at -20°C. Avoid repeated freeze/thaw. Keep on ice while in use. Stable for 2 months at -20°C.
- **Triglyceride Developer**: Reconstitute with 220 μl dH<sub>2</sub>O. Pipette up and down to dissolve completely. Aliquot and store at -20°C. Avoid repeated freeze/thaw. Stable for 2 months at -20°C.
- **Triglyceride Standard**: Storage at -20°C may cause the Triglyceride Standard to separate from the aqueous phase. To re-dissolve, tightly close the cap & put the vial in boiling water for 1 min (solution will turn cloudy). Vortex for 30 sec, the Standard solution will become clear. Repeat the heat and vortex step one more time. The Triglyceride Standard solution is now ready to be used.

## VIII. Triglyceride Assay Protocol:

**1. Sample Preparation:** Serum samples can be measured directly. Tissue (10 mg) or cells (1 x 10<sup>6</sup>) should be rapidly homogenized with 100 μl ice cold Triglyceride Assay Buffer for 10 min. on ice. Centrifuge at 12000 rpm for 5 min. Collect the supernatant. Saliva should be briefly spun down at 5000 rpm for 2 min. Collect the supernatant. Add 1-50 μl sample into a 96 well plate and adjust the volume to 100 μl/well with Triglyceride Assay Buffer.

### Notes:

- For unknown samples, we suggest testing several doses to ensure the readings are within the Standard Curve range.
- NADH in samples will generate background. For samples having high NADH levels, prepare parallel sample well(s) as background control.

