

ChromaFlash™ One-Step ChIP Kit

Base Catalog # P-2025

PLEASE READ THIS ENTIRE USER GUIDE BEFORE USE

Uses: The ChromaFlash™ One-Step ChIP Kit is suitable for selective enrichment of a chromatin fraction containing specific DNA sequences in a high throughput format using chromatin isolated from various species, particularly mammals. Chromatin can be isolated by using your own successful method or, for your convenience and the best results, with Epigentek's ChromaFlash™ Chromatin Extraction Kit (Cat. # P-2001 for mammals, and Cat. # P-2022 for plants) optimized for use with this product. The target protein bound DNA prepared with the ChromaFlash™ One-Step ChIP Kit can be used for various downstream applications including PCR (ChIP-PCR), microarrays (ChIP-chip), and sequencing (ChIP-seq).

Use of EpiSonic™ Sonication: The ChromaFlash™ One-Step ChIP Kit is optimized for use with the EpiSonic™ Multi-Functional Bioprocessor 1000 (Cat. # EQC-1000) in order to speed up the ChIP process and increase enrichment efficiency. For use of this kit without the EpiSonic™, please see "Standalone Protocol".

Input Amount of Chromatin: The amount of chromatin for each reaction can be 0.1 µg (about 1×10^4 cells) to 15 µg (about 1.5×10^6 cells). For an optimal reaction, the input chromatin amount should be 5 to 10 µg (about 0.5 to 1×10^6 cells), as enrichment of target proteins to genome loci varies and some of the target proteins are of low abundance.

Starting Materials: Starting materials can include various tissue or cell samples such as cells from flask or microplate cultured cells, fresh and frozen tissues, etc.

Antibodies: Antibodies should be ChIP or IP grade as to recognize fixed and native proteins that are bound to DNA or other proteins. If you are using antibodies which have not been validated for ChIP, then appropriate control antibodies such as RNA Polymerase II (Cat. # A-2032) should be used to demonstrate that the antibody and chromatin are suitable for ChIP.

Internal Controls: Both negative and positive DNA controls are provided in this kit.

Precautions: To avoid cross-contamination, carefully pipette the sample or solution into the strip wells. Use aerosol-barrier pipette tips and always change pipette tips between liquid transfers. Wear gloves throughout the entire procedure. In case of contact between gloves and sample, change gloves immediately.

KIT CONTENTS

Component	48 reactions Cat. #P-2025-48	96 reactions Cat. #P-2025-96	Storage Upon Receipt
CH1 (10X Wash Buffer)	10 ml	20 ml	4°C
CH2 (ChIP Buffer)	6 ml	12 ml	RT
CH3 (DNA Release Buffer)	14 ml	28 ml	RT
Non-Immune IgG (1 mg/ml)*	10 µl	20 µl	4°C
Anti-RNA Polymerase II (1 mg/ml)*	8 µl	16 µl	4°C
Proteinase K (10 mg/ml)*	55 µl	110 µl	4°C
GAPDH Primer - Forward (20 µM)*	8 µl	16 µl	4°C
GAPDH Primer - Reverse (20 µM)*	8 µl	16 µl	4°C
8-Well Assay Strips (With 1 Frame)	6	12	4°C
8-Well Strip Caps	6	12	RT
Sonication Frame	1	1	RT
96-Well PCR Plate	1	1	RT
Adhesive Covering Film	1	2	RT
User Guide	1	1	RT

* Spin the solution down to the bottom prior to use.

SHIPPING & STORAGE

The kit is shipped in two parts: the first part at ambient room temperature and the second part on frozen ice packs at 4°C.

Upon receipt: (1) Store **CH1, Non-Immune IgG, Anti-RNA Polymerase II, Proteinase K, GAPDH Primer – Forward, GAPDH Primer – Reverse, and 8-Well Assay Strips (With 1 Frame)** at 4°C away from light; (2) Store remaining components at room temperature away from light.

All components of the kit are stable for 6 months from the date of shipment, when stored properly.

Note; Check if **CH1** (10X Wash Buffer) contains salt precipitates before use. If so, briefly warm at room temperature or 37°C and shake the buffer until the salts are re-dissolved.

MATERIALS REQUIRED BUT NOT SUPPLIED

- Variable temperature waterbath or incubator oven
- Thermalcycler with 48 or 96-well block
- EpiSonic™ Multi-Functional Bioprocessor 1000 (Epigentek, Cat. # EQC-1000)
- Incubator oven with variable temperature

- Adjustable pipette and multiple-channel pipette
- Aerosol resistant pipette tips
- 0.2 ml or 0.5 ml PCR vials
- Antibodies of interest
- Orbital shaker (for standalone protocol)

GENERAL PRODUCT INFORMATION

Quality Control: Each lot of the ChromaFlash™ One-Step ChIP Kit is tested against predetermined specifications to ensure consistent product quality. Epigentek guarantees the performance of all products in the manner described in our product instructions.

Product Warranty: If this product does not meet your expectations, simply contact our technical support unit or your regional distributor. We also encourage you to contact us if you have any suggestions about product performance or new applications and techniques.

Safety: Suitable lab coat, disposable gloves, and proper eye protection are required when working with this product.

Product Updates: Epigentek reserves the right to change or modify any product to enhance its performance and design. The information in this User Guide is subject to change at any time without notice. Thus, only use the User Guide that was supplied with the kit when using that kit.

Usage Limitation: The ChromaFlash™ One-Step ChIP Kit is for research use only and is not intended for diagnostic or therapeutic application.

Intellectual Property: The ChromaFlash™ One-Step ChIP Kit and methods of use contain proprietary technologies by Epigentek.

A BRIEF OVERVIEW

Protein-DNA interaction plays a critical role for cellular functions such as signal transduction, gene transcription, chromosome segregation, DNA replication and recombination, and epigenetic silencing. Identifying the genetic targets of DNA binding proteins and knowing the mechanisms of protein-DNA interaction is important for understanding cellular process.

Chromatin immunoprecipitation (ChIP) offers an advantageous tool for studying protein-DNA interactions. It allows for the detection that a specific protein binds to the specific sequences of a gene in living cells by PCR (ChIP-PCR), microarrays (ChIP-chip), or sequencing (ChIP-seq). For example, measurement of the amount of methylated histone H3 at lysine 9 (meH3-K9) associated with a specific gene promoter region under various conditions can be achieved through a ChIP-PCR assay, while the recruitment of meH3-K9 to the promoters on a genome-wide scale can be detected by ChIP-chip. In particular, ChIP with antibodies directly against various transcriptional factors is widely demanded.

However, currently used ChIP methods have several drawbacks of which the most critical weakness is lengthy procedures, often taking up to 3 days to finish the procedures. Additionally, the labor-intensive procedure involves an excessive amount of steps, inconsistency, and sub-optimized chromatin shearing. These flaws result in inconvenience, low throughput processing, and less enrichment

efficiency. Epigentek first addressed these issues in 2005 by reducing the entire ChIP procedure to 5 hours in addition to improving upon performance and efficiency, and now further refines its ChIP expertise with the ChromaFlash™ technology.

Because the major features of next generation sequencing and microarrays are their rapidness and high throughput capabilities, these technologies are becoming major players in massive protein-DNA analysis. To be compatible with these new technologies, rapid and massive generation of target protein-bound DNA is critically required. To meet this requirement, Epigentek further developed a new ChIP technology called, ChromaFlash™ and incorporated it into its ChromaFlash™ One-Step ChIP Kit. With this kit and utilizing the EpiSonic™ Multi-Functional Bioprocessor 1000, chromatin shearing and immunoprecipitation can be simultaneously processed, which greatly advances ChIP to the highest speeds in a high throughput format with higher efficiency.


The ChromaFlash™ One-Step ChIP Kit has the following advantages:

- *The fastest and most convenient ChIP method.* The entire procedure (from intact chromatin sample to ready-for-use DNA) is less than 60 minutes with the actual handling time being less than 10 minutes due to simultaneous processing of chromatin shearing and immunoprecipitation ("One-Step ChIP").
- *96-well plate format makes the assay flexible.* Either (a) manual with one single reaction each time; or (b) high throughput with 96 reactions each time.
- *Highly efficient enrichment.* Enrichment ratio of positive to negative control > 120, and an extremely low number of cells required (as low as 10,000 cells per ChIP reaction).
- *High reproducibility.* Pre-optimized ChIP conditions and— with the EpiSonic Multi-Functional Bioprocessor 1000— digitally acoustic-controlled reaction processing in sealed vials make the ChIP procedure consistent.
- *Wide downstream analysis compatibility.* Compatible with various downstream analysis workflows including ChIP-PCR, ChIP-on-chip, and ChIP-seq.

PRINCIPLE & PROCEDURE

The ChromaFlash™ One-Step ChIP Kit contains all necessary reagents required for carrying out a successful chromatin immunoprecipitation directly from chromatin extracts isolated from mammalian cells or tissues. This kit includes a positive control antibody (RNA polymerase II), a negative control non-immune IgG, and GAPDH primers that can be used as a positive control to demonstrate the efficacy of the kit reagents and protocol. RNA polymerase II is considered to be enriched in the GAPDH gene promoter that is expected to be undergoing transcription in most growing mammalian cells and can be immunoprecipitated by RNA polymerase II but not by non-immune IgG. Immunoprecipitated DNA is then cleaned, released, and eluted. Eluted DNA can be used for various downstream applications such as ChIP-PCR, ChIP-on-chip, and ChIP-seq.





Schematic procedure of the ChromaFlash™ One-Step ChIP Kit




Fig 1. The data above shows the analysis of enrichment of RNA polymerase II in GAPDH and MLH1 promoters by the ChromaFlash™ One-Step ChIP Kit, with chromatin extract prepared from formaldehyde fixed colon cancer cells. Captured DNA was used for analyzing levels of RNA polymerase II enriched in the GAPDH and MLH1 promoters.

ASSAY PROTOCOL

For the best results, please read the protocol in its entirety prior to starting your experiment.

Starting Materials

Input Chromatin Amount: Chromatin amount can range from 0.1 µg to 15 µg per reaction. An optimal amount is 5-10 µg per reaction.

Chromatin Isolation: You can use your method of choice for chromatin isolation. Epigentek offers the ChromaFlash™ Chromatin Extraction Kit (Cat. # P-2001) for your convenience.

Chromatin Storage: Isolated chromatin can be stored at -20°C (short term) or -80°C (long term) until use.

1. Preparation of 1X Wash Buffer (CH1)

48 Reactions Kit: Add 10 ml of **CH1** 10X Wash Buffer to 90 ml of distilled water (pH 7.2-7.5).

96 Reactions Kit: Add 20 ml of **CH1** 10X Wash Buffer to 180 ml of distilled water (pH 7.2-7.5).

This **Diluted CH1** 1X Wash Buffer can now be stored at 4°C for up to six months.

2. Preparation of One-Step ChIP Reaction

- a. Predetermine the number of strip wells required for your experiment. Carefully remove the strip wells you need from the plate frame and transfer them to the reverse side of the **Sonication Frame** (upside down). Unused strip wells can be placed back in the bag (seal the bag tightly and store at 4°C).
- b. Setup the one-step ChIP reactions by adding the reagents to each well according to the following:

Reagents	Sample	Positive Control	Negative Control
CH2 ChIP Buffer	50-60 μ l	50-60 μ l	50-60 μ l
Chromatin	40-50 μ l	40-50 μ l	40-50 μ l
Your Antibodies	0.5-2 μ l	0	0
RNA Polymerase II	0	0.8 μ l	0
Non-Immune IgG	0	0	0.8 μ l

Note: The final amount of each component should be (a) chromatin: 5-10 μ g/well; (b) antibodies of interest: 0.8 μ g/well; (c) RNA Polymerase II: 0.8 μ g/well; and (d) non-immune IgG: 0.8 μ g/well.

Freshly prepared chromatin can be used directly for the reaction. Frozen chromatin samples should be thawed quickly at RT and then placed on ice before use. Store remaining chromatin samples at -20°C , or at -80°C if they will not be used within 8 hours.

The amounts of the positive control and negative control are sufficient for matched use with samples if two antibodies are used for each sample or one antibody is used for two of the same samples. If using one antibody of interest for each sample with matched use of the positive and negative control, extra RNA polymerase II and non-immune IgG required can be separately obtained from Epigentek.

Input DNA control is only used for estimating the enrichment efficiency of ChIP and is generally not necessary as the positive and negative control can be used for estimating the same objective more accurately.

*If you would like to include the input DNA control, the following steps can be carried out: (1) add 10 μ l of each chromatin sample to a 0.2 ml PCR tube followed by adding 88 μ l of **CH3** and 2.5 μ l of Proteinase K; (2) incubate the input DNA control at 65°C for 15 min, then incubate at 95°C for 10 min; and (3) spin the solution down to the bottom. Input DNA is ready for PCR or storage at -20°C .*

- c. Seal the wells with the **Adhesive Covering Film** and place the wells held on the **Sonication Frame** to the platform of the sample processing horn of the EpiSonic™ Multi-Functional Bioprocessor 1000.
- d. Start up the EpiSonic™ by following the standard EpiSonic™ operation instructions. Slowly add ice-cold filtered water into the sample processing horn so that your sample contents in the strip wells are submerged just below the water level.
- e. Set the EpiSonic™ to the following program using the power output chart under "Optimization Suggestions" in its manual as a reference:

20 sec Pulse-ON at 120W
 20 sec Pulse-OFF
 Total 20 duty cycles (Total ON time should be 3:40)
 10 min ON at 40W

For processing ChIP without using the EpiSonic™, see “Standalone Protocol.”

- f. Gently remove the **Sonication Frame** without tilting and carefully peel away the **Adhesive Covering Film** to avoid contamination between each well.

3. Washing of the Reaction Wells

- a. Carefully remove the solution and discard from each well.
- b. Wash each well with 200 µl of the **Diluted CH1** each time for four times. This can be done by simply pipetting **Diluted CH1** in and out of the well.
- c. Wash each well with 200 µl of **CH3** one time by pipetting **CH3** in and out of the well.

4. Reversal of Cross-Links, Release and Elution of DNA

- a. Prepare the **CH3-PK solution** by adding 1 µl of **Proteinase K** to each 39 µl of **CH3**, and mix.
- b. Add 40 µl of the **CH3-PK solution** to each well, then cover with strip cap.
- c. Incubate the wells at 65°C for 15-20 min.
- d. Quickly transfer the DNA solution from each well to 0.2 ml strip PCR tubes. Cap the PCR tubes.
- e. Incubate the PCR tubes containing DNA solution at 95°C for 5-10 min in a thermalcycler.
- f. Place the PCR tubes in room temperature. If liquid is collected on the inside of the caps, briefly spin the liquid down to the bottom.

DNA is now ready for use or storage at –20°C.

For real time PCR analysis, we recommend the use of 1-2 µl of eluted DNA in a 20 µl PCR reaction. If input DNA will be used, it should be diluted 10 fold before adding to PCR reaction. Control primers (110 bp, for human cells) included in the kit can be used as a positive control. For end point PCR, the number of PCR cycles may need to be optimized for better PCR results.

In general, the amplification difference between “normal IgG control” and “positive control” may vary from 3 to 8 cycles, depending on experimental conditions.

For ChIP-chip or ChIP-seq, additional DNA clean up and concentration steps may be needed. For your convenience, Epigentek offers a DNA Concentrator Kit (Cat. # P-1006) for DNA clean up and concentration.

TROUBLESHOOTING

Problem	Possible Cause	Suggestion
Little or No PCR Products Generated from both Sample and Positive Control Wells	Poor chromatin quality due to insufficient amount of cells, or insufficient or over cross-linking.	To obtain an optimal amount of chromatin per ChIP reaction should be 0.5-10 µg (about 0.5-1 x 10 ⁶ cells). The minimum amount of chromatin is 0.05 µg (5,000 cells).

		Appropriate chromatin cross-linking is also required. Insufficient or over-crosslinking will cause DNA loss or increased background. During cross-linking step of chromatin preparation, ensure that the cross-linking time is within 10-15 min, the concentration of formaldehyde is 1% as the final concentration, and/or quench solution is 1.25 M glycine.
	Poor enrichment with antibody; some antibodies used in ChIP might not efficiently recognize fixed protein.	Increase the antibody amount and use ChIP-grade antibodies validated for use in ChIP.
	Inappropriate DNA fragmenting condition.	If chromatin is from specific cell/tissue types such as plant, or is differently fixed, the processing program must be modified (see EpiSonic™ manual) to optimize the processing results. If using a probe-based sonicator, shearing conditions should also be optimized to allow DNA fragment size to be between 200-1000 bp.
	Incorrect temperature and/or insufficient time during DNA release.	Ensure the incubation times and temperatures described in the protocol are followed correctly.
	Improper PCR conditions, including improper PCR programming, PCR reaction solutions, and/or primers.	Ensure the PCR is properly programmed. If using a homebrew PCR reaction solution, check if each component is correctly mixed. If using a PCR commercial kit, check if it is suitable for your PCR. Confirm species specificity of primers. Primers should be designed to cover a short sequence region (70-150 bp) for more efficient and precise amplification of target DNA region (binding sites of the protein of interest).
	Inproper sample storage.	Chromatin sample should be stored at -80°C for no longer than 6 months, preferably less than 3 months. Avoid repeated freeze/thaw cycles. DNA samples should be stored at -20°C for no longer than 6 months, preferably less than 3 months.
No Difference in Signal Intensity Between	Insufficient washing.	Check if washing recommendations at each step is performed according to the

Negative and Positive Control Wells		<p>protocol. If the signal intensity in the negative control is still high, washing stringency can be increased in the following ways:</p> <ol style="list-style-type: none"> 1. Increase wash time at each wash step: after adding diluted CH1, leave it in the tubes/wells for 2-3 min and then remove it. 2. Add an additional one to two washes: The provided volume of Diluted CH1 is sufficient for 4 extra washes for each sample.
	<p>Too many PCR cycles; Plateau phase of amplification caused by over-increased number of PCR cycles in endpoint PCR may mask the difference of signal intensity between negative control and positive control.</p>	<p>Decrease the number of PCR cycles (i.e., 32-35 cycles) to keep amplification at the exponential phase will reduce high background in endpoint PCR and allow differences in amplification to be seen.</p> <p>Real time PCR is another choice in such cases.</p>
Little or No PCR Products Generated From Sample Wells Only	<p>Poor enrichment with antibody; some antibodies used in ChIP might not efficiently recognize fixed protein.</p>	<p>Increase the antibody amount and use ChIP-grade antibodies validated for use in ChIP.</p>
	<p>PCR primers are not optimized.</p>	<p>Confirm species specificity of primers. Primers should be designed to cover a short sequence region (70-150 bp) for more efficient and precise amplification of target DNA region (binding sites of the protein of interest).</p>

STANDALONE PROTOCOL

This protocol is intended for use without the EpiSonic™ Multi-Functional Bioprocessor 1000. For the best results, please read the protocol in its entirety prior to starting your experiment.

Starting Materials

Input Chromatin Amount: Chromatin amount can range from 0.1 µg to 15 µg per reaction. An optimal amount is 5-10 µg per reaction.

Chromatin Isolation: You can use your method of choice for chromatin isolation. Epigenetek offers the ChromaFlash™ Chromatin Extraction Kit (Cat. # P-2001) for your convenience.

Chromatin Storage: Isolated chromatin can be stored at -20°C (short term), or -80°C (long term) until use.

1. Chromatin Shearing

If a probe-based sonicator will be used, the sonication settings need to be optimized by you. For example, DNA of 200-1000 bp size can be obtained by sonicating 3-4 pulses of 10-12 sec each at level 2 using a Branson Microtip probe, followed by a 30-40 sec rest period on ice between each pulse. If desired, remove 10 μ l of sheared chromatin for DNA purification and agarose gel analysis along with a DNA marker on a 1 - 2% agarose gel, stained with ethidium bromide, and visualize it under ultraviolet light.

2. Preparation of 1X Wash Buffer (CH1)

48 Reactions Kit: Add 10 ml of **CH1** 10X Wash Buffer to 90 ml of distilled water (pH 7.2-7.5).

96 Reactions Kit: Add 20 ml of **CH1** 10X Wash Buffer to 180 ml of distilled water (pH 7.2-7.5).

This **Diluted CH1** 1X Wash Buffer can now be stored at 4°C for up to six months.

3. Preparation of One-Step ChIP Reaction

- Predetermine the number of strip wells required for your experiment. Carefully remove un-needed strip wells from the plate frame and place them back in the bag (seal the bag tightly and store at 4°C).
- Setup the ChIP reactions by adding the reagents to each well according to the following chart:

Reagents	Sample	Positive Control	Negative Control
CH2 ChIP Buffer	50-60 μ l	50-60 μ l	50-60 μ l
Chromatin	40-50 μ l	40-50 μ l	40-50 μ l
Your Antibodies	0.5-2 μ l	0	0
RNA Polymerase II	0	0.8 μ l	0
Non-Immune IgG	0	0	0.8 μ l

Note: The final amount of each component should be (a) chromatin: 5-10 μ g/well; (b) antibodies of interest: 0.8 μ g/well; (c) RNA Polymerase II: 0.8 μ g/well; and (d) non-immune IgG: 0.8 μ g/well.

Freshly prepared chromatin can be directly used for the reaction. Frozen chromatin samples should be thawed quickly at RT and then placed on ice before use. Store remaining chromatin samples at -20°C, or at -80°C if they will be not used within 8 hours.

The amounts of the positive control and negative control are sufficient for matched use with samples if two antibodies used for each sample or one antibody is used for two of the same samples. If using one antibody of interest for each sample with matched use of the positive and negative control, extra RNA polymerase II and non-immune IgG required can be separately obtained from Epigenetek.

Input DNA control is only used for estimating the enrichment efficiency of ChIP and is generally not necessary as the positive and negative control can be used for estimating the same objective more accurately.

*If you would like to include the input DNA control, the following steps can be carried out: (1) add 10 μ l of each chromatin sample to a 0.2 ml PCR tube followed by adding 88 μ l of **CH3** and 2.5 μ l of Proteinase K; (2) incubate the input DNA control at 65°C for 15 min followed by incubating at 95°C for 10 min; and (3) spin the solution down to the bottom. Input DNA is ready for PCR or storage at -20°C.*

- c. Seal the wells with **Adhesive Covering Film** and incubate the wells at room temperature for 90-120 min on an orbital shaker (100 rpm).
- d. Peel away the **Adhesive Covering Film** carefully to avoid contamination between each well.

4. Washing of the Reaction Wells

- a. Carefully remove the solution and discard from each well.
- b. Wash each well with 200 μ l of the **Diluted CH1** each time for four times. This can be done by simply pipetting **Diluted CH1** in and out of the well.
- c. Wash each well with 200 μ l of the **CH3** one time by pipetting **CH3** in and out of the well.

5. Reversal of Cross-Links, Release and Elution of DNA

- a. Prepare **CH3-PK** solution by add 1 μ l of **Proteinase K** to each 39 μ l of **CH3** and mix.
- b. Add 40 μ l of the **CH3-PK** solution to each well, then cover with strip cap.
- c. Incubate the wells at 65°C for 15-20 min.
- d. Quickly transfer the DNA solution from each well to 0.2 ml strip PCR tubes. Cap the PCR tubes.
- e. Incubate the PCR tubes containing DNA solution at 95°C for 5-10 min in a thermalcycler .
- f. Place the PCR tubes in room temperature. If liquid is collected on the inside of the caps, briefly spin the liquid down to the bottom.

DNA is now ready for use or storage at -20°C .

For real time PCR analysis, we recommend the use of 1-2 μ l of eluted DNA in a 20 μ l PCR reaction. If input DNA will be used, it should be diluted 10 fold before adding to PCR reaction. Control primers (110 bp, for human cells) included in the kit can be used as a positive control. For end point PCR, the number of PCR cycles may need to be optimized for better PCR results.

In general, the amplification difference between “normal IgG control” and “positive control” may vary from 3 to 8 cycles, depending on experimental conditions.

For ChIP-chip or ChIP-seq, additional DNA clean up and concentration steps may be required. For your convenience, Epigenetek offers a DNA Concentrator Kit (Cat. # P-1006) for DNA clean up and concentration.

PCR ANALYSIS

Real Time PCR

Primer Design

Primers designed should meet the criteria for real time PCR. For example, the covered sequence region should be 50-150 bp in length. G/C stretches at 3' ends of primers should be avoided.

PCR Reaction

Real time PCR can be performed using your own proven method. For your convenience and best

results, Epigentek offers the EpiQuik™ Quantitative PCR Kit (Cat #P-1029) which is optimized for fast qPCR reactions. As an example, the protocol is presented below:

Prepare the PCR Reactions

Thaw all reaction components including master mix, DNA/RNA free water, primer solution and DNA template. Mix well by vortexing briefly. Keep components on ice while in use, and return to -20°C immediately following use. Add components into each well according to the following:

Component	Size (µl)	Final Concentration
Methylamp Master Mix (2X)	10 µl	1X
Forward Primer	1 µl	0.4-0.5 µM
Reverse Primer	1 µl	0.4-0.5 µM
DNA Template	1-2 µl	50 pg-0.1 µg
DNA/RNA-free H ₂ O	6-7 µl	
Total Volume	20 µl	

For the negative control, use DNA/RNA-free water instead of DNA template.

Program the PCR Reactions

Place the reaction plate in the instrument and set the PCR conditions as follows:

Cycle Step	Temp	Time	Cycle
Activation	95°C	7 min	1
Cycling	95°C	10 sec	40
	55°C	10 sec	
	72°C	8 sec	
Final Extension	72°	1 min	1

Fold Enrichment Calculation

Fold enrichment (FE) can be calculated by simply using a ratio of amplification efficiency of the ChIP sample over that of non-immune IgG. Amplification efficiency of Polymerase RNA II can be used as a positive control.

$$FE \% = 2^{(IgG\ CT - Sample\ CT)} \times 100\%$$

For example, if CT for IgG is 38 and the sample is 34, then...

$$FE \% = 2^{(38 - 34)} \times 100\% = 1600\%$$

Endpoint PCR

Primer Design

Primers designed should meet the criteria for endpoint PCR. For example, the covered sequence region should be 100-400 bp in length. PCR primer design tools (e.g., Primer3Plus) can be used to help in the selection of appropriate primer pairs.

PCR Reaction

Endpoint PCR can be performed using your own proven method. It is important to stop the PCR reaction at the exponential phase by setting up an appropriate number of PCR cycles in order to make

a reliable comparison of enrichment efficiency obtained from different CHIP reactions. Thus, the optimized number of PCR cycles should be determined empirically.

PCR Product Analysis

Endpoint PCR products can be analyzed by separating amplicons on a 1-2% agarose gel, followed by staining with ethidium bromide and visualizing with UV-illumination.

RELATED PRODUCTS

CHIP Reaction

P-2026 ChromaFlash™ Magnetic One-Step CHIP Kit

Chromatin Preparation and Cleanup

P-2001 ChromaFlash™ Chromatin Extraction Kit

P-2023 ChromaFlash™ Chromatin Isolation and Shearing Kit

P-1006 DNA Concentrator Kit

Sonication Instruments

EQC-2000 EpiSonic™ 2000 Sonication System

PCR Analysis

P-1029 EpiQuik™ Quantitative PCR Kit

For CHIP-Grade Antibodies, search "chip-grade" at www.epigentek.com

